OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: January	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
2	Sim	MI 3351900	(1) .082	G0101
		Certification Number: 22096	(2)	Signature: B. Kelly /03
9	Sim	MI 3351900	(1) .081	G0101
		Certification Number: 22096	(2)	Signature: B. Kelly /03
12	Operation: Have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Geer, Instrument Manufacturer's Authorized Representative. 120 Day	MI	(1) .080	Ampoule Control Number: G0201 Guth .080 lot 09190 exp 6-15-10
		Certification Number: 22096	(2)	Signature: Marvin Geer
15	Sim	MI 3351900	(1) .078	G0101
		Certification Number: 11419	(2)	Signature: Donald R #147
19	Sim	MI 3351900	(1) .081	G0101
		Certification Number: 22096	(2)	Signature: B. Kelly /03
29	Sim	MI 3351900	(1) .080	G0101
		Certification Number: 22096	(2)	Signature: B. Kelly /03
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Supervisor Signature | Date

Accurate test results shall be .076 to .084 inclusive.

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month Feb / 02	Year 2010
Instrument Location LPD	
Instrument Number 202072	ORI Number MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
9	Sim	MI 3351900	(1) .080	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
18	Sim	MI 3351900	(1) .078	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
27	Sim	MI 3351900	(1) .077	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature | Date

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: March	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
6	Sim	MI 3351900	(1) .077	G0101
		Certification Number: 18861	(2)	Signature: #145
16	Sim	MI 3351900	(1) .080	G0101
		Certification Number: 22096	(2)	Signature: B. Kell 03
25	Sim	MI 3351900	(1) .078	G0101
		Certification Number: 22096	(2)	Signature: B. Kell 03
29	Sim	MI 3351900	(1) .079	G0101
		Certification Number: 18861	(2)	Signature: #145
Day	Operation	Operator ORI Number: MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature | Date

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
APRIL	2010

Instrument Location	
LPD	

Instrument Number	ORI Number
202072	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
4	SIM	MI 3351900	(1) .077	G020
		Certification Number 11419	(2)	Signature Donald Pt
12	SIM	MI 3351900	(1) .080	G0201
		Certification Number 11419	(2)	Signature Donald Pt #147
21	SIM	MI 3351900	(1) .079	G0201
		Certification Number 11419	(2)	Signature Donald Pt 2
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dahlke #397

Date 10-10-10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: May	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
5/6/10	Sim.	MI 3351900	(1) .076	G0201
		Certification Number: 19733	(2)	Signature: [signature] #95
10	Operation: [illegible] serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier, Instrument Manufacturer's Authorized Representative. 120 Day	MI /	(1) .079	G0201 Guth. [illegible] 09190 exp 6-15-10
		Certification Number: 99998	(2) /	Signature: Marvin Gier
14	Simulator	MI 3351900	(1) .079	G0201
		Certification Number: 10636	(2)	Signature: Dat W. Dahlke #397
20	Simulator	MI 3351900	(1) .078	G0201
		Certification Number: 10636	(2)	Signature: Dat W. Dahlke
23	Sim	MI 3351900	(1) .078	G0201
		Certification Number: 22096	(2)	Signature: B. [illegible] 03
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 6-2-10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
June	2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
6-3	Simulator	MI 3351900	(1) .076	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
6-9	Simulator	MI 3351900	(1) .081	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
6-15	Simulator	MI 3351900	(1) .080	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
6-24	Simulator	MI 3351900	(1) .079	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
6-30	Simulator	MI 3351900	(1) .078	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dahlke #397 Date 7-7-10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: July	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
7-7-10	Simulator	MI 3351900	(1) .083	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
7-13-10	Simulator	MI 3351900	(1) .080	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
7-21-10	Simulator	MI 3351900	(1) .080	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 10-10.10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: AUGUST	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
10	SIMULATION	MI 3351900	11919	.077		G0201	#147
15	SIM	MI 3351900	22096	.079		G0201	53
28	SIMULATION	MI 3351900	11919	.076		G0201	
Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
		MI					
Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
		MI					
Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
		MI					
Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
		MI					
Day	Operation	Operator ORI Number	Certification Number	Results (1)	Results (2)	Ampoule Control Number	Signature
		MI					

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Schlke #397

Date: 10-10.10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: September Year 2010
Instrument Location: LPD
Instrument Number: 202072
ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
02	I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier 120 Day Instrument Manufacturer's Authorized Representative	MI /	(1) .079	H0701 Gutt-0806 to 09340 exp 12-2-11
		Certification Number 99998	(2) /	Signature Marvin Gier
9-7-10	Simulator	MI 3351900	(1) .081	G0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-17-10	Simulator	MI 3351900	(1) —	H0201
		Certification Number 10636	(2) .080	Signature D. Dahlke #397
9-21-10	Simulator	MI 3351900	(1) .079	H0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-26-10	Simulator	MI 335190	(1) .080	H0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397
Date: 10-10-10

OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month: Oct.	Year: 2010
Instrument Location: LPD	
Instrument Number: 202072	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
1	Sim	MI 3351900	(1) .080	H0201
		Certification Number: 22096	(2)	Signature: B. Kell[illegible] 03
Day 10-3-10	Operation: Simulator	Operator ORI Number: MI 3351900	Results (1) .081	Ampoule Control Number: H0201
		Certification Number: 10636	(2)	Signature: D. Dahlke #397
Day 10-18-10	Operation: Simulator	Operator ORI Number: MI 3351900	Results (1) .080	Ampoule Control Number: F0801
		Certification Number: 10636	(2)	Signature: D. Dahlke #397
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
		Supervisor Signature		Date

Accurate test results shall be .076 to .084 inclusive.