OD-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	January
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
2	Sim	MI 3351900	(1) .081	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
9	Sim	MI 3351900	(1) .081	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
12	I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order Marvin Gyer Instrument Manufacturer's Authorized Representative 120 Day	MI	(1) .080	Ampoule Control Number G0101 Guth .080 Lot 09190, exp 6-19-10
		Certification Number 99998	(2) .081	Signature Marvin Gyer
15	Sim	MI 3351900	(1) .078	G101
		Certification Number 11419	(2)	Signature Donald [illegible] #147 11419
19	Sim	MI 3351900	(1) .082	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
29	Sim	MI 3351900	(1) .081	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature

Date

D-33 (7-03)
ICHIGAN STATE POLICE
PECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Feb / 02
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
9	Sim	MI 3351900	(1) .080	G0101
		Certification Number 22096	(2)	Signature B. Kelly 03
18	Sim	MI 3351900	(1) .079	G0101
		Certification Number 22096	(2)	Signature B. Kelly 03
27	Sim	MI 3351900	(1) .078	G0101
		Certification Number 22096	(2)	Signature B. Kelly 03
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature

Supervisor Signature

Date

Accurate test results shall be .076 to .084 inclusive.

D-33 (7-03)
IICHIGAN STATE POLICE
PECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	March
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
6	Sim	MI 3351900	(1) .078	G0101
		Certification Number 18861	(2) —	Signature [signature]
12	Taken out of service. Service call put in.	MI 3351900	(1)	
		Certification Number 22096	(2)	Signature B. Kelly /03
15	Operation: I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier, Instrument Manufacturer's Authorized Representative. Service key board locked up / Time adjusted	MI —	(1) .079	G0201 Guth 080609190 exp 06-15-2010
		Certification Number 99998	(2) —	Signature Marvin Gier
16	Sim	MI 3351900	(1) .081	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
25	Sim	MI 3351900	(1) .078	G0101
		Certification Number 22096	(2)	Signature B. Kelly /03
29	Sim	MI 3351900	(1) .079	G0101
		Certification Number 18861	(2)	Signature [signature] #145
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature

Supervisor Signature

Date

Accurate test results shall be .076 to .084 inclusive.

D-33 (7-03)
ICHIGAN STATE POLICE
PECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	APRIL
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
4	SIM	MI 3351900	(1) .076	G0201
		Certification Number 11419	(2)	Signature Donald Porter
12	SIM	MI 3351900	(1) .080	G0207
		Certification Number 11419	(2)	Signature Donald Pt
21	SIM	MI 3351900	(1) .078	G0201
		Certification Number 11419	(2)	Signature Donald Pt
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dahlke #397

Date 10-18-10

ID-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	May
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
5/6/10	Sim	MI 3351900	(1) .077	G0201
		Certification Number 19733	(2)	Signature #95
10	Operation: I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier 120 Day Instrument Manufacturer's Authorized Representative	MI /	(1) .080	G0201 Guth-080104 09190 exp 6-15-10
		Certification Number 99998	(2) /	Signature Marvin Gier
14	Simulator	MI 3351900	(1) .080	G0201
		Certification Number 10636	(2)	Signature D. Dale #397
23	Sim	MI 3351900	(1) .079	G0201
		Certification Number 22096	(2)	Signature B. Kelly 03
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dale #3

Date 6-2-10

ID-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	June
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
03	Simulator	MI 3351900	(1) .078	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
08	Simulator	MI 3351900	(1) .082	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
15	Simulator	MI 3351900	(1) .082	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
24	Simulator	MI 3351900	(1) .080	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
30	Simulator	MI 3351900	(1) .079	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Supervisor Signature D. Dahlke #397

Date 7-7-10

Accurate test results shall be .076 to .084 inclusive.

ID-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	July
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
7-7-10	Simulator	MI 3351900	(1) .084	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number	Results	Ampoule Control Number
7-13-10	Simulator	MI 3351900	(1) .082	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number	Results	Ampoule Control Number
7-21-10	Simulator	MI 3351900	(1) .082	G0201
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dahlke #397

Date 10-3-10

ID-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
Aug	2010
Instrument Location: LPD	
Instrument Number: 931024	ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
10	Simulator	MI 3351900	(1) .079	G0201
		Certification Number: 11419	(2)	Signature: Donald [illegible] #147
15	Sim	MI 3351900	(1) .080	G0201
		Certification Number: 22096	(2)	Signature: B. Kelly [illegible] 03
28	Simulator	MI 3351900	(1) .078	G0201
		Certification Number: 11419	(2)	Signature: Donald [illegible]
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 10-3-10

ID-33 (7-03)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	September
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
02	I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier 120 Day Instrument Manufacturer's Authorized Representative	MI —	(1) .078	40701 Guth .080 Lot 09340 exp 12-2-11
		Certification Number 99998	(2) —	Signature Marvin Gier
9-7-10	Simulator	MI 3351900	(1) .082	G0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-17-10	Simulator	MI 3351900	(1) .082	H0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-21-10	Simulator	MI 3351900	(1) .082	H0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-26-10	Simulator	MI 3351900	(1) .080	H0201
		Certification Number 10636	(2) —	Signature D. Dahlke #397
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature D. Dahlke #397

Date 10-3-10

D-33 (7-03)
IICHIGAN STATE POLICE
PECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Oct
Year	2010
Instrument Location	LPD
Instrument Number	931024
ORI Number	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
1	Sim	MI 3351900	(1) .083	H0201
		Certification Number 22096	(2)	Signature B. Kelly #03
30	Simulator	MI 3351900	(1) .082	H0201
		Certification Number 10636	(2)	Signature
10-10	Simulator	MI 3351900	(1) .082	F0801
		Certification Number 10636	(2)	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Supervisor Signature

Date

Accurate test results shall be .076 to .084 inclusive.